



NC DEPARTMENT OF
**HEALTH AND
HUMAN SERVICES**
Division of Public Health

ROY COOPER • Governor

MANDY COHEN, MD, MPH • Secretary

DANNY STALEY • Director, Division of Public Health

SCOTT J. ZIMMERMAN, DrPH, MPH, HCLD (ABB) • Laboratory
Director, State Laboratory of Public Health

Memo

To: Evelyn Foust
Communicable Disease Branch Head

From: Angela Bradley, MLT (ASCP)
Medical Laboratory Technologist

CC: Pete Moore
Ron Higginbotham
NCSLPH Lab Improvement Staff
Local Health Department Lab Managers

Date: September 18th, 2018

According to the NC State Laboratory of Public Health quality control evaluation, the following lot number(s) are **accepted** for use:

Bio-Bag™ Type C: Lot Number: 8129674 Expiration Date: 01/31/2020

Summary of NCSLPH QC results for this Lot #:

<i>N. gonorrhoeae</i> ATCC# 43069 - Growth
--

<i>N. gonorrhoeae</i> ATCC# 43070 - Growth
--

Lot number failed: None

BD and Fisher Scientific contacted: September 13th, 2018.

*If your contact information changes, please provide that information to Angie Bradley in Laboratory Improvement at (919) 733-7186 or via e-mail at angela.bradley@dhhs.nc.gov. Thank you.

NC DEPARTMENT OF HEALTH AND HUMAN SERVICES • DIVISION OF PUBLIC HEALTH • STATE LABORATORY OF PUBLIC HEALTH

LOCATION: 4312 District Drive, Raleigh, NC 27607

MAILING ADDRESS: 1918 Mail Service Center, Raleigh, NC 27699-1918

www.ncdhhs.gov • <http://publichealth.nc.gov> • slph.ncpublichealth.com • TEL: 919-733-7834 • FAX: 919-733-8695

AN EQUAL OPPORTUNITY / AFFIRMATIVE ACTION EMPLOYER